


Virus Like Particle on SRT SEC-1000

Column: SRT SEC-1000 7.8 x 300 mm, Flow rate: 1 mL/min, Detection: UV 280 nm,
 Mobile phase: 25 mM Tris, 150 mM NaCl, 1 mM CaCl₂ pH 7.5, Injection Volume: 50 µL
 VLP (14 million Da) 1mg/mL, Pressure: 51 bar




#	Time	Area	Height	Width	Area%	Symmetry
1	5.595	13.8	2.3E-1	0.9976	0.445	0.145
2	7.638	1692.7	74.6	0.3488	54.702	0.893
3	8.87	54.6	2	0.4019	1.763	1.303
4	9.664	76.3	5.9	0.2052	2.466	1.203
5	10.182	388.1	20.6	0.2757	12.542	0.684
6	10.797	768.3	24.9	0.4397	24.828	1.57
7	11.168	100.7	7	0.2151	3.254	0.279



Virus Like Particle on SRT SEC-2000

Column: SRT SEC-2000 7.8 x 300 mm, Flow rate: 1 mL/min, Detection: UV 280 nm, Mobile phase: 25 mM Tris, 150 mM NaCl, 1 mM CaCl₂, Injection Volume: 50 μ L VLP (14 million Da) 1mg/mL, Pressure: 52 bar



#	Time	Area	Height	Width	Area%	Symmetry
1	8.646	1668.2	84.4	0.3067	48.381	0.969
2	9.562	34.7	1.9	0.294	1.008	1.434
3	10.136	125.8	10.7	0.1849	3.649	3.608
4	10.466	570.7	33.7	0.2473	16.552	1.841
5	10.798	1048.6	37.1	0.392	30.410	0.774



VLP separation on SRT SEC-1000 and SEC-2000 in Tandem

Column: SRT SEC-2000 + SRT SEC-1000 7.8 x 300 mm, Flow rate: 1 mL/min, Detection: UV 280 nm, Mobile phase: 25 mM Tris, 150 mM NaCl, 1 mM CaCl₂, Injection Volume: 50 μ L VLP (14 million Da) 1mg/mL, Pressure: 52 bar

