

Compound extraction from whole lung of rats with Precellys®24 Dual - 7 ml vial.

Chiesi Farmaceutici, DMPK department - Italy

CONTEXT

The Company is focused on respiratory diseases. Our goal is the detection of chemicals entities as potential news drugs in biological fluids and tissues, as lung, bronchus and trachea after administration to animal

The sample preparation involves homogenization of tissues, most of the times, whole organ hence the trials with Precellys®24 Dual – 7 ml vial.

MATERIAL

- -Precellys®24 Dual vs Ultraturrax (usual device)
- -Precellys® kit 7ML ceramic bead CK28 (réf.0904-01)
- -Samples: 8 whole lungs weights ranged from 1.2 to 1.4 g (4 samples by device)
- -Buffer: Acetonitril / 0.9%NaCl mixture (50/50 v/v)

PROTOCOL

-Sampling: Whole lung (1 vol.) + Buffer (3 vol.)
-Precellys®24 Dual parameters: 6500rpm, 2x20

sec. vs Ultraturrax : around 40s
-Analyze : Injection of 10 µl of supernatant
(after protein precipitation) in LC/MS/MS


system (API3000 Triple Quadrupole – Applied Biosystems).

Chiesi

NESULTS ₽ DISCUSSION

The analyte levels in the sample are almost the same, so no difference involving the two homogenizing techniques are detectable (See Figure 1). Reproducibility is good in both cases, taking into account the individual variability.

Figure 1 □ Ultraturrax ■ Precellys24 Dual



Few residual cartilaginous fragments (probably bronchus/trachea) were observed in both methods (fewer amounts with Ultraturrax). The results show that there is no impact on the analyte recovery from matrix.

With both methods, temperature within the sample after processing could reach 70°C depending on the sample and protocols. Considering the stability of the analyte and the short time of exposure to this temperature (sample were preserved in ice before and after processing), sample degradation was not significant, but a refrigerating device could be useful in different situation.

Ultraturrax is fastidious and time consuming with cleaning steps and potential contamination between samples. The spent time is at least 5 times less using Precellys®24 Dual.

CONCLUSION

Precellys®24-Dual was successfully evaluated. As the results, **Precellys®24-Dual** is a suitable and reliable system **for grinding whole organs** with a fast process, without cross contamination, high reproducibility and comparable analyte levels extracted.

The use of **Cryolys cooling option** suits to reduce the increase temperature observed during processing.

Problem

Solution







For more details, please contact precellys@bertin.fr



ILIX機器株式会社

http://www.technosaurus.co.jp

■東京

〒162-0805 東京都新宿区矢来町 113番地 TEL(03)3235-0661(代) FAX(03)3235-0669

■大阪

〒532-0005 大阪市淀川区三国本町2丁目12番4号 TEL(06)6396-0501(代) FAX(06)6395-2588

■福岡

〒812-0054 福岡市東区馬出 1 丁目 2 番 23 号 TEL(092)631-1012(代) FAX(092)641-1285

※会社名および商品名は、各会社の商標または登録商標です。

※本カタログに記載の規格・仕様・外観は予告なく変更する場合がありますので御諒承下さい。