

# siRNA Analysis on Proteomix<sup>®</sup> NP5 SAX ( 4625 )

Column: Proteomix<sup>®</sup> NP5 SAX ( 5  $\mu$ m, 4.6 x 250 mm)

Mobile phase: A: 20 mM Tris, pH 8.0, B: A + 0.5 M NaClO<sub>4</sub>;

Gradient 3: 25-75% B in 30 min;

Flow rate: 0.5 mL/min; Detector: UV 260 nm, Injection volume: 2  $\mu$ L;

Column temperature: 25 °C; Pressure: 61 bar;

Samples: 2  $\mu$ L siRNA

